

Medical plaster bandages
 Made by PJSC «Ivano-Frankivskcement»

Technical characteristics		
Bandage characteristics	Nominal	Permissible variation
Lenth, (meter)	3.00 (2.70)	+ 0.05
Width, (millimeter)	200	± 5
Width, (millimeter)		
Width, (millimeter)		
Plaster setting time test on gauze, min.:		
- start	till 3	
- finish	by 8	
Bandage surface density, d/m ²	not less 300	
Plaster cementing surfacing, %	7	

Recommended loading rates		
Railway transport		Autotransport
Railway wagon	3300 boxes	Eurowagon 86 m ³ - 2800 boxes
20 feet railway container	1350 boxes	
5 t railway container	470 boxes	

Quantity of bandages in 1 box	
3.0x0.10 – 60 pcs.	2.7x0.10 – 68 pcs.
3.0x0.15 – 42 pcs.	2.7x0.15 – 45 pcs.
3.0x0.20 – 30 pcs.	2.7x0.20 – 34 pcs.

Manufacturer: PJSC «Ivano-Frankivskcement»
 Village Yamnytsa, Tysmenytsky district,
 Ivano-Frankivsk region
 Ukraine, 77422

E-mail: market@ifcem.if.ua
<http://www.ifcem.if.ua>

Tel.: +38 0342 58 36 67